## **BOSTER BIOLOGICAL TECHNOLOGY, LTD.**

3942 B Valley Ave, Pleasanton, CA, 94566 Phone: 925-485-4527. 888-466-3604 Fax: 925-485-4560

www.bosterbio.com

## **Boster Biotech Certificate of Analysis**

Note: this is a sample COA. To get the COA for your lot #, please contact us at support@bosterbio.com

To whom it may concern,

This letter attests that Human OPN / Osteopontin ELISA Kit ELISA Kit, catalog # EK0482, Lot # (this is SAMPLE COA) is manufactured by Boster Biological Technology, Ltd; through test and assay, we got the following data:


## 1. Typical Data

The following data were obtained for the various standards over the range of 156 pg/ml - 10000 pg/ml Human OPN / Osteopontin ELISA Kit , plus the blank well.

Concentrat	0	156	312	625	1250	2500	5000	10000
ion (pg/ml)								
O.D.	0.056	0.110	0.156	0.244	0.457	0.946	1.677	2.085

## **Standard Curve**





2. **Intra-Assay Precision** Samples of known Human OPN / Osteopontin ELISA Kit concentration were assayed in replicates of 16 to determine precision within an assay.

The %CV (Coefficient of Variation) of intra-assay is < 10%

3. **Inter-Assay Precision** Samples were assayed 48 times in multiple assays to determine precision between assays.

The %CV (Coefficient of Variation) of inter-assay is < 10%

- 4. **Sensitivity** The minimum detectable dose of Human OPN / Osteopontin ELISA Kit is<10 pg/ml. This was determined by adding two standard deviations to the mean O.D. Obtained when the zero standard was assayed 30 times.
- 5. **Expiration Date** Since the manufacture date for this kit is May 13, 2016, so, it can be stored at 4 degree for 6 months (November 13, 2016), and at -20 degree for 8 months (May 13, 2017). Note: these dates are demo data and not corresponding to the lot you are going to purchase. To get the COA for your particular lot, please contact us at support@bosterbi.com

Approved by:

Ms. Yangjing Yue Chief technician

Quality Control System of ELISA

Date: May 13, 2016