



Mouse Rat Triiodothyronine (T3) Total ELISA Kit

Catalog number: EK7011

For detection of multiple analytes using one single assay.

This package insert must be read in its entirety before using this product.

For research use only. Not for use in diagnostic procedures.

Mouse Rat Triiodothyronine (T3) Total ELISA Kit

Catalog Number: EK7011

Size: 96 wells/kit

Sample Type: Serum

Sensitivity: 0.25 ng/ml

Assay Range: 0.25-7.5 ng/ml

Storage: Store the kit at 2°C to 8°C. Keep microwells sealed in a dry bag with desiccants. The reagents are stable until expiration of the kit. Do not expose reagent to heat, sun, or strong light. Avoid multiple freeze-thaw cycles (Ships with gel ice, can store for up to 3 days in room temperature. Freeze upon receiving.)

Introduction

This ELISA kit is of competitive format. Competitive ELISA, also known as inhibition ELISA, is a surface/plate based assay, where the plate is coated with capture antibodies reactive to the molecule of interest. The sample (containing native molecule of interest) and enzyme conjugated recombinant protein (the competing molecule) are added to the coated wells. Since the amount of enzyme conjugated molecule in each well is constant, the level of native molecule in the sample will determine the binding ratio of enzyme conjugated molecule vs. native molecule. After an incubation period, any unbound antibody is washed off. Enzyme substrate (for example, TMB for HRP) is added to each well and will be transformed into a blue precipitate, the amount of which is linearly proportional to the amount of enzyme in the well. The precipitate is then turned into yellow by adding the acid stop solution and the concentration of yellow precipitate is read at 450nm for light absorbance (O.D. value). The O.D. is then used to calculate the amount of molecule of interest in each well, by comparing each sample well against the standard curve. The standard curve is generated using the same principle but instead of adding samples, a series of recombinant molecules with known concentrations are added to 6-8 wells.

Kit Components

Description	Quantity
Microwells coated with T3 Monoclonal Ab	12x8x1 Microwells
T3 Standards	7 vials (ready to use) 0.25 ml
T3 Enzyme Conjugate concentrate	1 vial 1.5ml
Assay Diluent	1 bottle (ready to use) 12ml
TMB Substrate	1 bottle (ready to use) 12ml
Stop Solution	1 bottle (ready to use) 12ml
20X Wash concentrate	1 bottle 25ml

Standard Concentrations and example data

	OD 450 nm	Conc. ng/mL
Std 1	2.404	0
Std 2	2.238	0.25
Std 3	2.001	0.5
Std 4	1.714	1
Std 5	1.147	2.5

Std 6	0.793	5
Std 7	0.642	7.5

Materials Required, but Not Provided

1. Distilled or deionized water
2. Precision pipettes
3. Disposable pipette tips. Multichannel pipettes are recommended in the condition of large amount of samples in the detection.
4. ELISA reader capable of reading absorbance at 450nm
5. Absorbance paper or paper towel
6. Graph paper

WARNINGS AND PRECAUTIONS

1. For Research Use Only. Not for use in diagnostic procedures.
2. Potential biohazardous materials: The calibrator and controls contain human source components, which have been tested and found non-reactive for hepatitis B surface antigen as well as HIV antibody with FDA licensed reagents. However, there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent. These reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories" 1984.
3. Do not pipette by mouth. Do not smoke, eat, or drink in the areas in which specimens or kit reagents are handled.
4. The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed.
5. It is recommended that standards, control and serum samples be run in duplicate.
6. Optimal results will be obtained by strict adherence to this protocol. Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential. Any deviation from this may yield invalid data.
7. Do not use **sodium azide** as preservative. Sodium azide inhibits HRP enzyme activities.

SPECIMEN COLLECTION HANDLING

1. Collect blood specimens and separate the serum immediately.
2. Typically, specimens may be stored refrigerated at (2°C to 8°C) for 5 days. If storage time exceeds 5 days, store frozen at (-20°C) for up to one month.
3. Avoid multiple freeze-thaw cycles.

4. Prior to assay, frozen sera should be completely thawed and mixed well.

5. Do not use grossly lipemic specimens.

REAGENT PREPARATION

T3-enzyme Conjugate Solution: Dilute the T3-enzyme conjugate 1:11 with assay diluent in a suitable container. For example, dilute 160µl of conjugate with 1.6ml of assay diluent for 16 wells (A slight excess of solution is made). This reagent should be used within twenty-four hours for maximum performance of the assay. Store at 2-8°C.

General Formula:

Amount of Buffer required = Number of wells * 0.1

Quantity of T3-Enzyme necessary = # of wells * 0.01

i.e. = 16 x 0.1 = 1.6ml for Assay diluent

16 x 0.01 = 0.16ml (160 ul) for T3 enzyme conjugate.

Wash Buffer: Prepare 1X Wash buffer by adding the contents of the bottle (25 ml, 20X) to 475 ml of distilled or deionized water. Store at room temperature (20-25°C).

ASSAY PROCEDURE

Before proceeding with the assay, bring all reagents, serum references and controls to room temperature (20-25°C).

1. Format the microplate wells for each serum reference, control and patient specimen to be assayed in duplicate. Replace any unused microwell strips back into the aluminum bag, seal and store at 2-8°C.
2. Pipette 25 ul of the appropriate serum reference, control or specimen into the assigned well.
3. Add 100 ul of working T3-enzyme conjugate solution to all wells (see Reagent Preparation Section).
4. Cover the plate and Incubate for 60 minutes at room temperature with shaking at 650rpm.
5. Remove liquid from all wells. Wash wells three times with 300 of 1X wash buffer (see Reagent Preparation Section). Blot on absorbent paper towels.
6. Add 100 ul of TMB substrate solution to all wells.
7. Cover the plate and Incubate at room temperature for fifteen (15) minutes.
8. Add 50 ul of stop solution to each well and gently mix for 15-20 seconds.
9. Read the absorbance on ELISA Reader of each well at 450nm within 15 minutes after adding the stop solution.

CALCULATION OF RESULTS

The standard curve is constructed as follows:

1. Check T3 standard value on each standard vial. This value might vary from lot to lot. Make sure you check the value on every kit. See example of the standard attached.
2. To construct the standard curve, plot the absorbance for T3 standards (vertical axis) versus T3 standard concentrations (horizontal axis) on a linear graph paper. Draw the best curve through the points.
3. Read the absorbance for controls and each unknown sample from the curve. Record the value for each control or unknown sample.

Submit a Product Review to Biocompare.com

Submit a review of this product to Biocompare.com to receive a \$20 Amazon.com gift card! Your reviews help your fellow scientists make the right decisions. Thank you for your contribution.



Mouse Rat Triiodothyronine (T3) Total ELISA Kit