



PicoKine™ Quick ELISA Kit

Catalog number: FEK0577

For the quantitation of **Mouse Il6st** concentrations in cell culture supernatants, serum and plasma (heparin, EDTA, citrate).

This package insert must be read in its entirety before using this product.
For research use only. Not for use in diagnostic procedures.

Mouse Gp130/IL6ST PicoKine® Quick ELISA Kit

Catalog Number:FEK0577

Assay Principle

The Boster Picokine™ Mouse Il6st Pre-Coated ELISA (Enzyme-Linked Immunosorbent Assay) kit is a solid phase immunoassay specially designed to measure Mouse Il6st with a 96-well strip plate that is pre-coated with antibody specific for Il6st. The detection antibody is a biotinylated antibody specific for Il6st. The capture antibody is monoclonal antibody from rat and the detection antibody is polyclonal antibody from goat. The kit includes Mouse Il6st protein as standards.

To measure Mouse Il6st, add standards and samples to the wells, then add the biotinylated detection antibody. Wash the wells with PBS or TBS buffer, and add Avidin-Biotin-Peroxidase Complex (ABC-HRP). Wash away the unbounded ABC-HRP with PBS or TBS buffer and add TMB. TMB is substrate to HRP and will be catalyzed to produce a blue color product, which changes into yellow after adding acidic stop solution. The density of the yellow product is linearly proportional to Mouse Il6st in the sample. Read the density of the yellow product in each well using a plate reader, and benchmark the sample wells' readings against the standard curve to determine the concentration of Mouse Il6st in the sample. For more information on assay principle, protocols, and troubleshooting tips, see Boster's ELISA Resource Center at <https://www.bosterbio.com/elisa-technical-resource-center>.

Overview

Product Name	Mouse Gp130/IL6ST PicoKine® Quick ELISA Kit
Reactive Species	Mouse
Size	96 wells/kit, with removable strips.
Description	Mouse Gp130/IL6ST PicoKine™ Quick ELISA Kit (90 minutes, 96 Tests). Quantitate Mouse Il6st in cell culture supernatants, serum and plasma (heparin, EDTA, citrate). Sensitivity: 10pg/ml.
Sensitivity	< 10 pg/ml *The sensitivity or the minimum detectable dose (MDD) is the lower limit of target protein that can be detected by the kit. It is determined by adding two standard deviations to the mean O.D. value of twenty (20) blank wells and calculating the corresponding concentration.
Detection Range	125 pg/ml - 8,000 pg/ml
Storage Instructions	Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles (Ships with gel ice, can store for up to 3 days in room temperature. Freeze upon receiving.)
Uniprot ID	Q00560

Technical Details

Capture/Detection Antibodies	The capture antibody is monoclonal antibody from rat and the detection antibody is polyclonal antibody from goat.
Specificity	Natural and recombinant Mouse Il6st

Standard	Expression system for standard: NS0; Immunogen sequence: Q23-E617
Cross Reactivity	There is no detectable cross-reactivity with other relevant proteins.

Kit Components/Materials Provided

Description	Quantity	Volume
Anti-Mouse Il6st Pre-coated 96-well strip microplate	1	12 strips of 8 wells
Mouse Il6st Standard	2	10 ng/tube
Mouse Il6st Biotinylated antibody (50x)	1	100 µl
Avidin-Biotin-Peroxidase Complex (30x)	1	400 µl
Sample Diluent	1	30 ml
Antibody Diluent	1	12 ml
Avidin-Biotin-Peroxidase Diluent	1	12 ml
Wash Buffer (25x)	1	20 ml
Color Developing Reagent (TMB)	1	10 ml
Stop Solution	1	10 ml
Plate Sealers	4	Piece

Required Materials That Are Not Supplied

Microplate Reader capable of reading absorbance at 450nm.

Automated plate washer (optional)

Pipettes and pipette tips capable of precisely dispensing 0.5 µl through 1 ml volumes of aqueous solutions.

Multichannel pipettes are recommended for large amount of samples.

Deionized or distilled water.

500ml graduated cylinders.

Test tubes for dilution.

Notice Before Application

Please read the following instructions before starting the experiment.

1. To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.
2. Before using the Kit, spin tubes and bring down all components to the bottom of tubes.

3. Don't let 96-well plate dry, for dry plate will inactivate active components on plate.
4. Don't reuse tips and tubes to avoid cross contamination.
5. Avoid using the reagents from different batches together.


Mouse Gp130/IL6ST PicoKine® Quick ELISA Kit (FEK0577) Standard Curve Example

Highest O.D. value might be higher or lower than in the example. The experiment result is statistically significant if the highest O.D. value is no less than 1.0.

Concentration (pg/ml)	0	125	250	500	1000	2000	4000	8000
O.D.	0.088	0.133	0.199	0.307	0.484	0.829	1.254	1.911

Mouse Gp130/IL6ST PicoKine ELISA Kit standard curve

A standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.



Intra/Inter Assay Variability

Boster spend great efforts in documenting lot to lot variability and make sure our assay kits produce robust data that are reproducible.

Intra-Assay Precision (Precision within an assay): Three samples of known concentration were tested on one plate to assess intra-assay precision.

Inter-Assay Precision (Precision accross assays): Three samples of known concentration were tested in separate assays to assess inter-assay precision.

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	16	16	16	24	24	24
Mean (pg/ml)	159	1404	4588	158	1473	4382
Standard deviation	11.6	71.6	339.51	14.06	81.01	333.03
CV (%)	7.3%	7.6%	7.4%	8.9%	5.5%	7.6%

Reproducibility

We ensure reproducibility by testing three samples with differing concentrations of Bdnf in ELISA kits from four different production batches/lots.

Lots	Lot 1 (pg/ml)	Lot 2 (pg/ml)	Lot 3 (pg/ml)	Lot 4 (pg/ml)	Mean (pg/ml)	Standard Deviation	CV (%)
Sample 1	159	185	185	168	174	11.21	6.4%
Sample 2	1404	1352	1232	1273	1315	66.97	5%
Sample 3	4588	4716	4423	4674	4600	112.25	2.4%

*number of samples for each test n=16.

Preparation Before The Experiment

Item	Preparation
All reagents	Bring all reagents to 37°C prior to use. Also the TMB incubation time estimate (20-25min) is based on 37°C.
Wash buffer	Prepare 500 ml of Working Wash Buffer by diluting the supplied 20 ml of Wash Buffer (25 x) with 480 ml of deionized or distilled water. If crystals have formed in the concentrate, warm to room temperature and mix it gently until crystals have completely dissolved.
Biotinylated Anti-Mouse Il6st antibody	It is recommended to prepare this reagent immediately prior to use by diluting the Mouse Il6st Biotinylated antibody (50x) 1:50 with Antibody Diluent. Prepare 50 µl by adding 1 µl of Biotinylated antibody (50x) to 49 µl of Antibody Diluent. Mix gently and thoroughly and use within 2 hours of generation.
Avidin-Biotin-Peroxidase Complex	It is recommended to prepare this reagent immediately prior to use by diluting the Avidin-Biotin-Peroxidase Complex (30x) 1:30 with Avidin-Biotin-Peroxidase Diluent. Prepare 400 µl by adding 10 µl of Avidin-Biotin-Peroxidase Complex (30x) to 390 µl of Avidin-Biotin-Peroxidase Diluent. Mix gently and thoroughly and use within 2 hours of generation.
Mouse Il6st Standard	It is recommended that the standards be prepared no more than 2 hours prior to performing the experiment. Use one 10 ng of lyophilized Mouse Il6st standard for each experiment. Gently spin the vial prior to use. Reconstitute the standard to a stock concentration of 10 ng/ml using 1ml of sample diluent. Allow the standard to sit for a minimum of 10 minutes with gentle agitation prior to making dilutions.
Microplate	The included microplate is coated with capture antibodies and ready-to-use. It does not require additional washing or blocking. The unused well strips should be sealed and stored in the original packaging.

Dilution of Mouse Il6st Standard

1. Number tubes 1-8. Final Concentrations to be Tube # 1: 8,000.00 pg/ml, # 2: 4,000.00 pg/ml, # 3: 2,000.00 pg/ml, # 4: 1,000.00 pg/ml, # 5: 500.00 pg/ml, # 6: 250.00 pg/ml, # 7: 125.00 pg/ml, # 8: Sample Diluent serves as the zero standard (0 pg/ml).
2. To generate standard #1, add 800 µl of the reconstituted standard stock solution of 10 ng/ml and 200 µl of sample diluent to tube #1 for a final volume of 1000 µl. Mix thoroughly.
3. Add 300 µl of sample diluent to tubes # 2-7.
4. To generate standard # 2, add 300 µl of standard # 1 from tube # 1 to tube # 2 for a final volume of 600 µl. Mix thoroughly.
5. To generate standard # 3, add 300 µl of standard # 2 from tube # 2 to tube # 3 for a final volume of 600 µl. Mix thoroughly.
6. Continue the serial dilution for tube # 4-7.

Sample Preparation and Storage

These sample collection instructions and storage conditions are intended as a general guideline and the sample stability has not been evaluated.

Sample Type	Procedure
Cell culture supernatants	Clear sample of particulates by centrifugation, assay immediately, or store samples at -20°C.
Serum	Use a serum separator tube (SST) and allow serum to clot at room temperature for about four hours. Then, centrifuge for 15 min at approximately 1,000 x g. assay immediately or store samples at -20°C.
Plasma	Collect plasma using heparin, EDTA or citrate as an anticoagulant. Centrifuge for 15 min at approximately 1,000 x g. Assay immediately or store samples at -20°C. *Note: it is important to not use anticoagulants other than the ones described above to treat plasma, for other anticoagulants could block the antibody binding site.

Sample Dilution

The target protein concentration should be estimated and appropriate sample dilutions should be selected such that the final protein concentration lies near the middle of the linear dynamic range of the assay.

It is recommended to prepare 150 µl of sample for each replicate to be assayed. The samples should be diluted with sample diluent and mixed gently.

Assay protocol

It is recommended that all reagents and materials be equilibrated to 37°C/room temperature prior to the experiment (see Preparation Before The Experiment if you have missed this information).

1. Prepare all reagents and working standards as directed previously.
2. Remove excess microplate strips from the plate frame and seal and store them in the original packaging.
3. Add 50 µl of the standard, samples, or control per well. And add 50 µl of the prepared 1x Biotinylated Anti-Mouse IgG antibody per well. Add 50 µl of the sample diluent buffer and 50 µl of the prepared 1x Biotinylated Anti-Mouse IgG antibody into the control well (Zero well). At least two replicates of each standard, sample, or control is recommended.
4. Cover with the plate sealer provided and incubate for 60 minutes at RT.
5. Wash the plate 3 times with the 1x wash buffer.
 - a. Discard the liquid in the wells into an appropriate waste receptacle. Then, invert the plate on the benchtop onto a paper towel and tap the plate to gently blot any remaining liquid. It is recommended that the wells are not allowed to completely dry at any time.
 - b. Add 300 µl of the 1x wash buffer to each assay well. (For cleaner background incubate for 60 seconds between each wash).
 - c. Repeat steps a-b 2 additional times.
6. Add 100 µl of the prepared 1x Avidin-Biotin-Peroxidase Complex into each well. Cover with plate sealer provided and incubate for 15 minutes at RT.
7. Wash the plate 5 times with the 1x wash buffer.
 - a. Discard the liquid in the wells into an appropriate waste receptacle. Then, invert the plate on the benchtop onto a paper towel and tap the plate to gently blot any remaining liquid. It is recommended that the wells are not allowed to completely dry at any time.
 - b. Add 300 µl of the 1x wash buffer to each assay well. (For cleaner background incubate for 60 seconds between each wash).
 - c. Repeat steps a-b 4 additional times.

8. Add 90 µl of Color Developing Reagent to each well and incubate in the dark for 30 minutes at RT (or 25-30 minutes at 37°C). (The optimal incubation time must be empirically determined. A guideline to look for is blue shading the top four standard wells, while the remaining standards remain clear.)
9. Add 100 µl of Stop Solution to each well. The color should immediately change to yellow.
10. Within 30 minutes of stopping the reaction, the O.D. absorbance should be read with a microplate reader at 450nm.

Data Analysis

Average the duplicate readings for each standard, sample, and control. Subtract the average zero standard O.D. reading.

It is recommended that a standard curve be created using computer software to generate a four parameter logistic (4-PL) curve-fit. A free program capable of generating a four parameter logistic (4-PL) curve-fit can be found online at: www.myassays.com/four-parameter-logistic-curve.assay.

Alternatively, plot the mean absorbance for each standard against the concentration. The measured concentration in the sample can be interpolated by using linear regression of each average relative OD against the standard curve generated using curve fitting software. This will generate an adequate but less precise fit of the data.

For diluted samples, the concentration reading from the standard curve must be multiplied by the dilution factor.

Background on IL6st

GP130 (Glycoprotein 130) also known as IL6ST or CD130, is a transmembrane protein which is the founding member of the class of all cytokine receptors. The GP130 gene is mapped on 5q11.2. Gp130 showed an apparent molecular mass of 130 kD by SDS-PAGE. Gp130 is involved in the formation of high-affinity IL6-binding sites and in IL6 signal transduction. Sgp130 has a role in modulating signals transduced by membrane-bound gp130. Gp130 is an important part of many different types of signaling complexes. Gp130 molecules with mutations in either of these segments could not transduce a growth signal.

Submit a product review to Biocompare.com

Submit a review of this product to Biocompare.com to receive a \$20 Amazon giftcard! Your reviews help your fellow scientists make the right decisions. Thank you for your contribution.



Mouse Gp130/IL6ST ® Quick ELISA Kit