

## **BOSTER BIOLOGICAL TECHNOLOGY, LTD.**

3942 B Valley Ave, Pleasanton, CA, 94566 Phone: 925-485-4527. 888-466-3604 Fax: 925-485-4560

[www.bosterbio.com](http://www.bosterbio.com)

### **Boster Biotech Certificate of Analysis**

Note: this is a sample COA. To get the COA for your lot #, please contact us at [support@bosterbio.com](mailto:support@bosterbio.com)

To whom it may concern,


This letter attests that Human WISP1/CCN4 ELISA Kit, catalog # EK1204, Lot # (this is SAMPLE COA) is manufactured by Boster Biological Technology, Ltd; through test and assay, we got the following data:

#### **1. Typical Data**

The following data were obtained for the various standards over the range of 31.2pg/ml-2000pg/ml Human WISP1/CCN4 , plus the blank well.

Concentration (pg/ml)	0	31.2	62.5	125	250	500	1000	2000
O.D.	0.033	0.123	0.198	0.396	0.654	1.152	1.753	2.262

#### **Standard Curve**



2. **Intra-Assay Precision** Samples of known Human WISP1/CCN4 concentration were assayed in replicates of 16 to determine precision within an assay.

The %CV (Coefficient of Variation) of intra-assay is < 10%


3. **Inter-Assay Precision** Samples were assayed 48 times in multiple assays to determine precision between assays.

The %CV (Coefficient of Variation) of inter-assay is < 10%

4. **Sensitivity** The minimum detectable dose of Human WISP1/CCN4 is < 10pg/ml. This was determined by adding two standard deviations to the mean O.D. Obtained when the zero standard was assayed 30 times.

5. **Expiration Date** Since the manufacture date for this kit is May 13, 2016, so, it can be stored at 4 degree for 6 months (November 13, 2016), and at -20 degree for 8 months (May 13, 2017). Note: these dates are demo data and not corresponding to the lot you are going to purchase. To get the COA for your particular lot, please contact us at support@bosterbi.com

Approved by:



Ms. Yangjing Yue Chief technician

Quality Control System of ELISA

Date: May 13, 2016